



YOGI VEMANA UNIVERSITY

Department of Genetics and Genomics

Vemanapuram, Kadapa, AP-516005

Dr. L. Dakshayani
Assistant Professor

Mobile: 9652001184 Fax: 08562-225419

E-mail: dlomada@yahoo.com

No: DL-CMR/YVU/SERB /Equipment- Microtome/ Call for Quotations/2019

Date: 11-01-2019

QUOTATIONS REQUESTED

Sir,

Sub: SERB project sanctioned to Dr. L. Dakshayani, PI and Dr.C. Madhava Reddy, Co-PI- Purchase of Equipment - MICROTOME- Quotations requested -Reg.

Ref: SERB sanction orders file no. EMR/2016/007208, dated 17th Sep, 2018

You are requested to send your lowest quotation in duplicate towards the supply of the **MICROTOME** with specifications listed below so as to reach The Registrar, Yogi Vemana University, Vemanapuram, Kadapa, AP- 516 005 on or before 31/01/2019 by post/courier. The University is not responsible for the postal delays. The sealed quotation should invariably be super scribed as "Microtome No: DL-CMR/YVU/SERB /Equipment- Microtome/ Call for Quotations/2019 dated 11-01-2019". Quotations include, terms and conditions, unit price, taxes/duties, warranty period, validity of quotation, payment mode, transport charges, etc.

Equipment name	Technical specifications
Microtome (Made in Japan/ Germany/Singapore/ Indonesia/USA, Europe, Korea only, Please don't coat for domestic and made in China product)	<ul style="list-style-type: none"> • Manual Microtome with specimen feed system • Section thickness feed to the specimen from 0.5-60µm • Section thickness selection from 0.5–2 µm in 0.5 µm-steps • Trimming function in 4 steps • Hand wheel lock in each position • The object head with +/-8° X/Y orientation should be equipped with a fast specimen clamp exchange system for improved workflow • Must have maintenance-free vertical cross-roller guides and horizontal specimen feed via precision cylinder guide system. • Additional 12 O' Clock locking by simply pulling the handle • Vertical Cross roller bearings mechanism for maintenance free operation • Wide range of accessories- extra blade box, specimen clamp etc • CE / USFDA certificate • 2 KVA Stabilizer • Extended warranty for 3 years

Date: 11.01.2019

DR. L. DAKSHAYANI

PI, SERB. No. EMR/2016/007208
Dept. of Genetics & Genomics
YOGI VEMANA UNIVERSITY
Kadapa, AP-516005, India.